

Kalibrering af IR termometer m/type K stik 6 pkt.

Certificate of Calibration
Elma Instruments - Føttermarken 7 - DK - 3520 Faxe

Model: M8000 Element AT

PERFORMED ON: 08 Februar 2019

MANUFACTURER TYPE: Instruktionstype

TEST RESULT: PASS

MODEL NUMBER: 1130000

CERTIFICATE TYPE: F0200-001

PRODUCTION NAME: Model M8000 Element AT

TEMPERATURE: 22.00°C

ACCURACY: 0.5%

CERTIFICATE NO.: 29602552

CALIBRATED BY: Janus Sødermann

Elma Instruments A/S certifies that the above mentioned instrument meets or exceeds the manufacturer's published specifications at the points tested. All measurements are traceable to national and/or international standards, or have been derived by approved metrology labs.

This certificate may not be reproduced other than in full. Calibration certificates without signature are not valid.

Any Test Uncertainty Ratio (TUR) that is less than four to one will appear under the "TUR" heading on the data record. If the TUR meets or exceeds four to one, the field is left blank.

Measurements using 70% or more of the published specification is marked with a yellow box.

ITEMS:

Item No.	Description	S/N	Cal Date	Exp Date
001	Fluke 5700A Calibration	918720287	03.01.2019	03.01.2020

Test Results:

Test Description	True Value	Test Result	Lower Limit	Upper Limit	TUR	Status
Response						
Response 5.00 or above						
						Pass
Battery voltage						
						Pass
Insulation Resistance 1 M 250V						
0.00 MΩ	0.00	0.00	0.00	0.00		Pass
0.01 MΩ	0.00	0.00	0.00	0.00	195.00	Pass
0.02 MΩ	0.00	0.00	0.00	0.00	14.00	Pass
0.03 MΩ	0.00	0.00	0.00	0.00	40.00	Pass
Insulation Resistance 1 M 500V						
0.00 MΩ	0.00	0.00	0.00	0.00	270.00	Pass
0.01 MΩ	0.00	0.00	0.00	0.00	83.00	Pass

Issue Printed: August 07, 2018 Page 1 of 6

Sporbar Kalibrering

Har du styr på din målenøjagtighed?

Det er vigtigt, at du kan dokumentere nøjagtigheden på dit instrument.

Lad Elma Instruments udføre sporbar kalibrering – så er du sikker.

For at sikre, at dit instrument altid udfører rigtige og præcise målinger anbefaler vi, at det bliver kalibreret én gang om året. Når du indleverer et instrument til kalibrering hos Elma Instruments gennemgår vi følgende procedurer:

- Gennemgang og kontrol af hele instrumentet.
- Opdatering af firmware (instrument software).
- Kalibrering efter producentens anvisninger.
- Udstedelse af certifikat.
- Registrering af instrument og styring af genindkaldelse.

Du er altid velkomne til at kontakte os hvis du har spørgsmål.

E-mail: service@elma.dk

Telefon: 7022 1000 - Tryk 3

Tekniske Data: